

Welch Allyn Australia (Pty) Ltd. Unit 5, 38 – 46 South Street Rydalmere NSW 2116

## MANUFACTURER'S DECLARATION OF CONFORMITY AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002 DECLARATION OF CONFORMITY PROCEDURES

SAP DIR No.:

80020236

Version:

This is a declaration of conformity made under clause 6.6 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.

Manufacturer's name:

Welch Allyn, Inc.

Business address:

4341 State Street Road

Skaneateles Falls, NY 13153-0220

U.S.A.

Product name:

GS Lights

#

DEVICES:

44416, 44456, 44606, 44616, 44906, 44900-C, 44900-W, 48816

ACCESSORIES:

44215, 48200, 48605, 48805, 48850, 48950, 48955, 48960, 52630, 52640-B,

405966

Classification:

T

GMDN code and term:

12276 - Light, examination

Scope of application:

All

Each kind of medical device to which the technical documentation applies complies with the applicable provisions of the essential principles and the classification rules before being supplied.



Welch Allyn Australia (Pty) Ltd. Unit 5, 38 – 46 South Street Rydalmere NSW 2116

## MANUFACTURER'S DECLARATION OF CONFORMITY AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002 DECLARATION OF CONFORMITY PROCEDURES

DEEL MATTER OF CONTONINT TROCED CKES				
SAP DIR No.:	80020236	V	ersion:	A
Standards applied	d: EN 1	4971	2007	Medical Devices- Application of Risk Assessment to medical devices
		0601-1 (incl. ndments)	1990	Medical Electrical Equipment- part 1: General requirements for basic safety and essential requirements
	EN 6	0601-1-1	2000	Medical Electrical Equipment- part 1-1-General requirements for safety- Collateral Standard: Safety requirements for Medical Electrical Equipment
	EN 6	0601-1-2	2004	Medical Electrical Equipment- part 1-2- General requirements for safety- Collateral Standard: Electromagnetic Compability- Requirements and Test
	EN 6	0601-1-4	1997	Medical Electrical Equipment- part 1-4- General requirements for safety- Collateral Standard: General requirements for programmable electrical medical systems
	EN 6	0601-1-6	2004	Medical Electrical Equipment- part 1-6- General requirements for safety- Collateral Standard: Usability
	EN 10	041	2008	Information supplied by the manufacturer with Medical devices
	EN 9	80	2008	Graphical symbols for use in the labeling of medical devices
	EN 62	2366	2008	Medical devices- Application of usability engineering to medical devices
	ISO 1	4155	2003	Clinical investigation of medical devices for human subjects
	ISO 1	0993-1	2003	Biological evaluation of medical devices Part 1: Evaluation and testing

Authorised Signatory:

Tim Croft Sr. Manager, Regulatory Affairs - JAPAC Date Rydalmere, NSW Place of Issue

This authorisation is given in the signatory's capacity as representative of the "Manufacturer" (as recorded on page 1 of this declaration)